




Hip Pro Left Smart Wrap™ Quick Start Guide

Setting Up the Kelvi Pro^e Console





Applying the Smart Wrap™


- 4** Place the Smart Wrap™ on the left hip. Pull the waist strap **A** around the waist and through the D-ring. Secure the strap to itself.

The tube and connector should sit above the waist strap and exit towards the right side of the body.

- 5** Pull the thigh strap **B** under the leg and attach it to the main body of the Smart Wrap™. The position of the strap may be adjusted as needed for comfort.




Setting Up Treatment*



6 Select treatment type

On the touchscreen, select **Single Temperature** for constant treatment

Select **Contrast** for alternating cool and hot treatment




7 Set temperature and time

Select **Cool** or **Heat**, then select temperature.

Set a time limit or choose **Continuous**.

Press **Start** to begin treatment.

Smart Wrap™ may be adjusted as needed for comfort during treatment.



8 Adjust air pressure for fit

Select **+** or **-** to change air pressure.

Start low and increase by one level at a time. Do not exceed comfortable pressure.

Press **Stop** to release air pressure.

Press **Stop** to end treatment.

* Follow clinician prescription for treatment



Scan here
for complete
instructions
for use



IMPORTANT: Read complete instructions including indications, contraindications, cautions, and warnings before using this product. For assistance in using the Kelvi Pro™ System, contact your licensed or certified healthcare professional.

